

6-14-2017

A Primer to the Structure, Content and Linkage of the FDA's Manufacturer and User Facility Device Experience (MAUDE) Files

Lisa Garnsey Ensign

University of Colorado, Denver, Anschutz Medical Campus, lisa.ensign@ucdenver.edu

K. Bretonnel Cohen

University of Colorado, Denver, School of Medicine, kevin.cohen@gmail.com

Follow this and additional works at: <http://repository.edm-forum.org/egems>



Part of the [Analytical, Diagnostic and Therapeutic Techniques and Equipment Commons](#), [Biomedical Devices and Instrumentation Commons](#), [Databases and Information Systems Commons](#), [Health Information Technology Commons](#), and the [Health Services Research Commons](#)

Recommended Citation

Ensign, Lisa Garnsey and Cohen, K. Bretonnel (2017) "A Primer to the Structure, Content and Linkage of the FDA's Manufacturer and User Facility Device Experience (MAUDE) Files," *eGEMs (Generating Evidence & Methods to improve patient outcomes)*: Vol. 5: Iss. 1, Article 12.

DOI: <https://doi.org/10.13063/2327-9214.1253>

Available at: <http://repository.edm-forum.org/egems/vol5/iss1/12>

This Governance Review is brought to you for free and open access by the the Publish at EDM Forum Community. It has been peer-reviewed and accepted for publication in eGEMs (Generating Evidence & Methods to improve patient outcomes).

The Electronic Data Methods (EDM) Forum is supported by the Agency for Healthcare Research and Quality (AHRQ), Grant 1U18HS022789-01. eGEMs publications do not reflect the official views of AHRQ or the United States Department of Health and Human Services.

A Primer to the Structure, Content and Linkage of the FDA's Manufacturer and User Facility Device Experience (MAUDE) Files

Abstract

Introduction and Background: The US Food and Drug Administration (FDA)'s Manufacturer and User Facility Device Experience (MAUDE) database is a publicly available resource providing over 4 million records relating to medical device safety. Using downloadable MAUDE files avoids limitations of the online MAUDE search interface. However, naïve file usage can result in errors, while independent discovery of the nuances required to correctly work with the database can be time-consuming. Practical information is provided to shorten this learning curve and obtain accurate results when using the MAUDE database files.

MAUDE File Descriptions: The MAUDE database consists of 135 fields in 4 primary (Master Event, Device, Patient, Text) and 2 supplemental (Device Problems and Problem Code Descriptions) file types. When combined, these 6 files provide a detailed account of an adverse event or product problem report. Website instructions for joining the files are incomplete. Comprehensive details are provided to enable precise file linking.

Lessons Learned: MAUDE files have irregularities that must be understood to download and work with the data efficiently. Accurate results depend upon combining the files correctly and understanding the difference between report and event denominators. Appreciating data availability can facilitate successful MAUDE investigations.

Conclusion: The MAUDE database can provide key insights about medical device safety. Detailed information is provided about the structure, content and interrelationships of the MAUDE database files to enable investigators to use this valuable resource more quickly and accurately.

Acknowledgements

Supported by NIH/NCATS Colorado CTSI Grant Number UL1 TR001082. Contents are the authors' sole responsibility and do not necessarily represent official NIH views.

Keywords

Information Storage and Retrieval, Informatics, Health Information Technology, Data Use and Quality

Disciplines

Analytical, Diagnostic and Therapeutic Techniques and Equipment | Biomedical Devices and Instrumentation | Databases and Information Systems | Health Information Technology | Health Services Research

Creative Commons License



This work is licensed under a [Creative Commons Attribution-NonCommercial-No Derivative Works 3.0 License](https://creativecommons.org/licenses/by-nc-nd/3.0/).



A Primer to the Structure, Content and Linkage of the FDA's Manufacturer and User Facility Device Experience (MAUDE) Files

Lisa Garnsey Ensign,ⁱ K. Bretonnel Cohenⁱⁱ

ABSTRACT

Introduction and Background: The US Food and Drug Administration (FDA)'s Manufacturer and User Facility Device Experience (MAUDE) database is a publicly available resource providing over 4 million records relating to medical device safety. Using downloadable MAUDE files avoids limitations of the online MAUDE search interface. However, naïve file usage can result in errors, while independent discovery of the nuances required to correctly work with the database can be time-consuming. Practical information is provided to shorten this learning curve and obtain accurate results when using the MAUDE database files.

MAUDE File Descriptions: The MAUDE database consists of 135 fields in four primary (Master Event, Device, Patient, Text) and two supplemental (Device Problems and Problem Code Descriptions) file types. When combined, these six files provide a detailed account of an adverse event or product problem report. Website instructions for joining the files are incomplete. Comprehensive details are provided to enable precise file linking.

Lessons Learned: MAUDE files have irregularities that must be understood to download and work with the data efficiently. Accurate results depend upon combining the files correctly and understanding the difference between report and event denominators. Appreciating data availability can facilitate successful MAUDE investigations.

Conclusion: The MAUDE database can provide key insights about medical device safety. Detailed information is provided about the structure, content and interrelationships of the MAUDE database files to enable investigators to use this valuable resource more quickly and accurately.

ⁱUniversity of Colorado, Anschutz Medical Campus, ⁱⁱUniversity of Colorado, Denver, School of Medicine

Introduction

The increasing number of big data repositories and analytical tools available for medical research is one of the exciting trends of the 21st century.¹⁻³ The [US Food and Drug Administration \(FDA\)'s Manufacturer and User Facility Device Experience \(MAUDE\) database](#) is one such resource. MAUDE contains over four million medical device adverse event and product problem reports dating back to 1991. With nearly two thousand new adverse event and product problem reports submitted every day, the MAUDE database is an important tool for monitoring and investigating safety issues involving medical devices. MAUDE has facilitated the identification and investigation of medical product problems ranging from cardiovascular and gynecological devices to stretchers and tanning beds.⁴⁻¹³

While the FDA's online MAUDE search engine provides a targeted way to look up small numbers of adverse event reports associated with specific products or problems, the search results are limited to the past 10 years and detailed information is not structured in a format that can be readily analyzed. In addition, the number of records returned for a given query is capped, and in a small sample study nearly half of the search results exceeded this limit.

The search engine limitations may be overcome by using downloadable MAUDE files provided by the FDA. However, naïve linking of the six MAUDE file types that comprise an adverse event record can result in erroneous findings, while understanding the nuances of the various file types and how they work together can be exceedingly time-intensive.¹⁴⁻²¹

The MAUDE database contains medical device problem and outcome information that can inform general reviews, specific product inquiries, and examination of adverse event reporting trends over time. The goal of this paper is to provide novice

users with information gained through practical experience about the basic structure and content of the MAUDE database. Detailed information about the content, structure and technical issues of the files is also provided for more technical readers who need to obtain a complete picture of an adverse event report or an analysis database. The information detailed in this paper has the potential to shorten the learning curve and improve the accuracy of results obtained by researchers using MAUDE.

MAUDE Background

Medical devices encompass several thousand health care products, ranging from simple tongue depressors and bandages, to complex medical lasers and MRI machines. As defined by the FDA, a medical device is a product used to diagnose, treat, cure, mitigate, prevent or alter bodily functions. Although they may be combined with a drug as in the case of drug-eluting coronary stents, medical devices differ from pharmaceutical (drugs) or biological (blood, vaccine, gene or cellular) products in that they do not use metabolism, chemical or immunological means to achieve their primary intended purpose.²²

The MAUDE database contains adverse events and product problem reports involving medical devices, including reports of suspected device-associated deaths, serious injuries and malfunctions. The publicly available MAUDE database encompasses the releasable, medical device reporting information submitted through MedWatch Form 3500 (Form FDA 3500 A for Mandatory Reporting by Manufacturers and Form FDA 3500 for Voluntary Reporting by patients, health professional and consumers).²³ MAUDE also includes adverse event information submitted via the FDA's Alternative Summary Reporting (ASR) program.²⁴ Database records extend back to enactment of the Safe Medical Devices Act²⁵ and include reports submitted by user facilities since 1991, those from distributors



and voluntary submitters since 1993, and from manufacturers since 1996.²⁶ As shown in Figure 1, the vast majority of MAUDE reports are now submitted by manufacturers, and the number of submissions has grown exponentially over the last decade.


The Utility of MAUDE

The FDA uses MAUDE reports to monitor device performance, detect potential device-related safety issues, and inform the risk-benefit assessments of these products.^{27,28} Health care professionals use MAUDE to review events associated with specific products, body systems or procedures.²⁹⁻³⁴ More than 120 articles referencing MAUDE have been

published to date, the majority of these summarizing adverse events specific to a particular outcome, product or body system.

Two examples illustrate the potential of MAUDE: 1) In an investigation of complications associated with global endometrial ablation, Gurtcheff and Sharp⁸ identified injuries reported in MAUDE that were not yet identified in peer-reviewed publications, leading the authors to “encourage physicians to review the MAUDE database when considering the use of a new medical device, to research the possibility of complications not yet reported in the published medical literature.” 2) Hauser and colleagues³⁵ used

Figure 1. MAUDE Medical Device Reports through 2013 by Year Received and Reporting Source



information reported in the MAUDE database to investigate whether a new implantable cardioverter-defibrillator (ICD) lead coating effectively prevented insulation failures caused by abrasions. Based on their findings, the authors concluded that the material “did not prevent these abrasions” and raised awareness of a potential product issue subsequently observed by others.^{36,37} Finally, in our own research, we are using information in the MAUDE database to explore early predictors of medical device recall.³⁸

Limitations of Using MAUDE Search Results for Research Studies

The FDA provides access to MAUDE information through three mechanisms (Figure 2): an online simple (single-parameter)³⁹ or advanced (multi-parameter)⁴⁰ search interface or downloadable data files.²⁶ While the online search engines are extremely convenient, information obtained using these interfaces has several limitations:

1. Results are restricted to reports made within the past 10 years.
2. Only a small subset of MAUDE fields is provided in the comma-separated value file that may be downloaded from these results (Table 1).
3. Details for an individually-selected search result record may be viewed in a web browser, but the information is not provided in a structured manner that can be readily downloaded or analyzed (Figure 3).
4. Results are limited to 500 records per query which an inattentive user may miss (Figure 4) and which necessitates the use of narrowed search criteria, adding time to the process and potentially constraining the original question of interest.

The single-parameter (a) and multi-parameter (b) online MAUDE search interfaces, and an example of files available from the MAUDE download files interface (c).

This example illustrates a listing of MAUDE results obtained using the single-parameter MAUDE search interface where the query returned more than the maximum 500 records.

Using the advanced search interface, a small sample study was conducted to explore the extent to which search engine results exceeded the n=500 upper limit record restriction. This study examined the 79 Class 3 medical devices assigned for review by the cardiovascular device panel.⁴¹ One quarter (n=20) of the searches returned more than 500 results for 2013 alone, while 42 percent (n=33) of the same queries exceeded the count restriction for the 5-year period from 2009-2013. Often the easiest way to narrow the search results is to restrict the specified date interval; however, since the number of retrieved records is unknown, it can be quite time-consuming to iteratively determine date intervals that result in fewer than 500 records, while also minimizing the number of result files that must be separately exported and combined. For example, more than 500 coronary drug-eluting stents (Product Code NIQ) records were retrieved in 2013. Detailed iterations found that 2-3 week windows (equating to 16 date intervals) were necessary to optimize the number of results returned per year while staying below the 500 count maximum. Obviously, this manual process quickly becomes quite tedious if one wanted to revise the search to cover a longer time span or to investigate both drug- and non-drug eluting coronary stents.



Figure 2. Accessing MAUDE Data

(a)

Search Database Help Download Files

Date Report Received by FDA
(press CTRL key for multiple years)

- ALL YEARS
- 2016
- 2015
- 2014
- 2013

[Go to Advanced Search](#) Records per Report Page [Clear Form](#)

Enter a single word (e.g., electromechanical), an exact phrase (e.g., electromechanical pump) or multiple words. To Search by Brand Name, Manufacturer, Event Type, 510K Number, PMA Number, Product Code, or date, select [Go To Advanced Search](#) button.

(b)

Search Database Help Download Files

Product Problem

Product Class

Event Type Manufacturer

Model Number Report Number

Brand Name Product Code

Date Report Received by FDA (mm/dd/yyyy) to

[Go to Simple Search](#) Records per Report Page [Clear Form](#)

(c)

File Name	Compressed Size in Bytes	Uncompressed Size in Bytes	Total Records	
mdrfoithru2014.zip	142765KB	1232037KB	4083820	Master Record through 2014
patient.zip	2641KB	21044KB	744536	MAUDE Patient records received to date for 2015
patientadd.zip	155KB	1877KB	71223	New MAUDE Patient records for the current month.

The single-parameter (a) and multi-parameter (b) online MAUDE search interfaces, and an example of files available from the MAUDE download files interface (c).

Table 1. Only a Fraction of Data is Available when Exporting MAUDE Online Search Results to Excel

REPORT NUMBER	MANUFACTURER	BRAND NAME	DATE REPORT RECEIVED	PRODUCT CODE	EVENT DATE	EVENT TYPE	EVENT TEXT
Web Address: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/detail.cfm?mdrfoi_id=3545674							
2938836-10124	ST. JUDE MEDICAL; INC.; CRMD	RIATA ST OPTIM ACTIVE FIXATION	12/31/2013	LWS	10/4/2013	Malfunction	Event Description: IT WAS REPORTED THAT DURING DEVICE REPLACEMENT; LEAD DISLODGE MENT WAS NOTED. TESTS INDICATED THAT ALL THE ELECTRICAL PARAMETERS WERE GOOD AND THE PHYSICIAN ELECTED TO LEAVE THE LEAD IN ITS CURRENT POSITION. NO FURTHER ISSUES WERE REPORTED AND THE PATIENT CONDITION WAS GOOD AFTER THE EVENT. Manufacturer Narrative: ALL INFORMATION PROVIDED BY MANUFACTURER; NO MEDWATCH FORM WAS RECEIVED.
Web Address: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/detail.cfm?mdrfoi_id=3545673							
2938836-10118	ST. JUDE MEDICAL; INC.; CRMD	RIATA ACTIVE FIXATION	12/31/2013	LWS	8/12/2013	Malfunction	Manufacturer Narrative: ALL INFORMATION PROVIDED BY MANUFACTURER; NO MEDWATCH FORM WAS RECEIVED. Event Description: IT WAS REPORTED THAT DURING DEVICE CHANGE OUT DUE TO NORMAL ERI; MINOR INSULATION ABRASION WAS OBSERVED ON THE PROXIMAL PORTION OF THE LEAD. NO ELECTRICAL ANOMALIES WERE DETECTED. LEAD WAS REPAIRED WITH A LEAD REPAIR KIT AND REMAINS IMPLANTED. PATIENT WAS STABLE AND EXPERIENCED NO COMPLICATIONS DUE TO THE PROCEDURE.
Web Address: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/detail.cfm?mdrfoi_id=3545098							
2938836-10101	ST. JUDE MEDICAL; INC.; CRMD	RIATA ACTIVE FIXATION	12/31/2013	LWS	11/20/2013	Malfunction	Event Description IT WAS REPORTED THAT HIGH HV LEAD IMPEDANCE WAS OBSERVED VIA REMOTE TRANSMISSION. DEVICE WAS REPROGRAMMED AND IMPEDANCE MEASUREMENTS WERE NORMAL. A SUCCESSFUL INDUCTION TEST WAS PERFORMED. THE PATIENT CONDITION WAS GOOD AND WILL BE MONITORED. Manufacturer Narrative: ALL INFORMATION PROVIDED BY MANUFACTURER; NO MEDWATCH FORM WAS RECEIVED.



Figure 3. Detailed Information from the MAUDE Online Search Interface is Not Readily Analyzed

GE MEDICAL SYSTEMS, LLC SIGNA MR	
Model Number	2226300
Event Date	11/02/2006
Event Type	Injury
Event Description	Two days after an mr scan, a patient reported a small blister to their arm. The facility reported the padding had become dislodged in the area of the injury during the scan. According to the facility, the patient sustained blisters approximately 2 cm in diameter with the potential for scarring and treated the patient with antibiotics.
Manufacturer Narrative	System investigation is ongoing.
Brand Name	SIGNA
Type of Device	MR
Manufacturer (Section D)	GE MEDICAL SYSTEMS, LLC 3200 N. Grandview Blvd. Waukesha WI 53188
Manufacturer (Section G)	GE MEDICAL SYSTEMS, LLC 3200 N Grandview Blvd Waukesha WI 53188
Manufacturer Contact	Mary Overland, Ph. d., Manager 3000 N Grandview Blvd Waukesha , WI 53188 (262) 262 -2625 26254824 2625482402
MDR Report Key	805236
Report Number	2183553-2007-00001
Device Sequence Number	1
Product Code	LNH
Report Source	Manufacturer
Source Type	Health Professional,User facility
Reporter Occupation	Other
Type of Report	Initial
Report Date	01/12/2007
1 Device Was Involved in the Event	
1 Patient Was Involved in the Event	
Date FDA Received	01/12/2007
Is This An Adverse Event Report?	Yes
Is This A Product Problem Report?	No
Device Operator	Health Professional
Device MODEL Number	2226300
Was Device Available For Evaluation?	No
Is The Reporter A Health Professional?	Yes
Was the Report Sent to FDA?	No
Date Manufacturer Received	12/18/2006
Was Device Evaluated By Manufacturer?	No
Date Device Manufactured	10/01/2003
Is The Device Single Use?	No
Is this a Reprocessed and Reused Single-Use Device?	No
Type of Device Usage	Reuse
Patient TREATMENT DATA	
Date Received: 01/12/2007 Patient Sequence Number: 1	

Figure 4. Truncated Results with the MAUDE Online Search Interface

MAUDE - Manufacturer and User Facility Device Experience

FDA Home | Medical Devices | Databases

CDRH SuperSearch

DeNovo | Registration & Listing | Adverse Events | Recalls | PMA | HDE | Classification | Standards
CFR Title 21 | Radiation-Emitting Products | X-Ray Assembler | Medsun Reports | CLIA | TPLC

1 to 10 of 500 Results² for riata 2015

1 2 3 4 5 6 7 8 9 10 >

10 results per page

Manufacturer	Brand Name	Date Report Received
ST. JUDE MEDICAL, INC. (CRM-KISTA)	RIATA PASSIVE FIXATION	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-KISTA)	RIATA ACTIVE FIXATION	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ST OPTIM PASSIVE FIXATION	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ST OPTIM LEAD	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ST ACTIVE FIXATION	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-KISTA)	RIATA ST OPTIM PASSIVE FIXATION	12/30/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ACTIVE FIXATION	12/29/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ACTIVE FIXATION	12/29/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ST ACTIVE FIXATION	12/29/2015
ST. JUDE MEDICAL, INC. (CRM-SUNNYVALE)	RIATA ACTIVE FIXATION	12/29/2015

² The maximum 500 records meeting your search criteria returned. Please narrow your search.

This example illustrates a listing of MAUDE results obtained using the single-parameter MAUDE search interface where the query returned more than the maximum 500 records.

To avoid the time constraint and the field and count limitations of the Search interface, the FDA makes available downloadable MAUDE data sets (Figure 2c). While these files allow one to explore complex questions, to span large time frames and to retrieve detailed report information, the individual files must first be imported and pieced together. This process can be much more complex than initially anticipated due to the lack of detail and inconsistencies in the online information describing the MAUDE

downloadable files and their contents. The remainder of this paper focuses on documented tips and techniques for processing and combining these files based on empirically determined potential sources of error.

Organization of the MAUDE Files

As shown in Table 2, MAUDE files are grouped into four primary and two related supplemental files spanning a total of 135 fields:



Primary

- **Master Event:** the centerpiece to which all of the other MAUDE files eventually relate. Data in the Master Event file include adverse event and product problem information reported in sections B, E and H of the MedWatch Form 3500, along with detailed distributor and manufacturer contact and address information from MedWatch sections F and G.
- **Device:** includes baseline reporting information and information reported in section D of the MedWatch Form 3500, including device identifiers such as manufacturer, product name and model number.
- **Patient:** contains concatenated treatment and outcome information.
- **Text:** encompasses MedWatch Form 3500 narrative information from sections B and H describing the adverse event and the manufacturer's evaluation of returned devices.

The four primary file types provide options for selecting data for the current year, current month, previous year (Device and Text files) or years combined (Master Event and Patient data, and early Device (before 1998) and Text (before 1996) data, and file changes.

Supplemental

- **Device Problems:** links a MAUDE report to one or more device problem codes reported in section F of the MedWatch Form 3500.
- **Problem Code Descriptions:** maps a device problem code to a short, text description.

Given the current file structure, a Device and a Text file are added each year, and a complete MAUDE database covering information from 1991 to the present entails over 50 files.

MAUDE Files and Fields as defined on the MAUDE website. Added shading indicates file linking (see

also Figure 5). Information in parentheses indicates the MedWatch Form 3500 source field.

When combined, the information contained in the six MAUDE file types provide a detailed account of a reported adverse event or product problem report. However, while accurate, MAUDE website instructions indicating that "All record types are linked via the MDR REPORT KEY" are incomplete. Figure 5 illustrates the full interrelationships among the MAUDE data sets. The nuances of each file type and specific considerations for joining the files are reviewed in the next section.

Detailed MAUDE File Information

This section discusses specific information about importing and joining files together, as well as information about the content and structure of the files. In this discussion, MAUDE field names are displayed in italicized capital font, using the names provided on the MAUDE website.

Importing Files

The downloadable MAUDE files are provided as zipped files in which individual data elements are pipe-delimited. A header row containing field names is present for the Master Event and Device data sets, but labels must to be manually entered for the other file types. As such, when importing data, make sure the import process takes into account whether a header row is present.

While the MAUDE website indicates it may be necessary "to put in an extra character at the end of the first record prior to importing the file, otherwise the last column of data may be lost", in practice we did not find this step to be required when importing data into SAS 9 or RStudio. However, readers should be alert to the presence of rare, but potentially problematic, non-printing line-feed (LF) characters. Not only can these invisible characters


Table 2. MAUDE Files and Content

MASTER EVENT	DEVICE	PATIENT	TEXT	DEVICE PROBLEMS	PROBLEM CODE DESCRIPTIONS
Header Row: Yes	Header Row: Yes	Header Row: No	Header Row: No	Header Row: No	Header Row: No
File Names and Descriptions					
mdrfoi : Master Event records for the current year	foidev : Device data for the current year	patient : Patient data for the current year	foitext : Text data for the current year	foidevproblem : Device Problem codes	deviceproblemcode : Mapping of Problem Codes to their description
mdrfoiadd : New Master Event records for the current month	foidevadd : New Device data for the existing month	patientadd : New Patient data for the current month	foitextadd : New Text data for the existing month		
mdrfoichange : Changes to existing Master Event records	foidevchange : Changes to existing Device data	patientchange : Changes to existing Patient data	foitextchange : Changes to existing Text data		
mdrfoithruFFFF : Cumulative Master Event records, 1991-previous calendar year	foidevthru1997 : Cumulative Device data, 1991-1997	patientthruFFFF : Cumulative Patient data, 1991-previous calendar year	foitextthru1995 : Cumulative Text data, 1991-1995		
	foidevFFFF : Annual Device data, 1998-previous calendar year		foitextYYYY : Annual Text data, 1996-previous calendar year		
Fields					
1. MDR Report Key	46. Manufacturer Contact City (G1)	1. MDR Report Key	1. MDR Report Key	1. MDR Report Key	1. Device Problem Code (F10)
2. Event Key	47. Manufacturer Contact State Code (G1)	2. Device Event key	2. Patient Sequence Number	2. MDR Text Key	2. Problem Description
3. Report Number	48. Manufacturer Contact Zip Code (G1)	3. Implant Flag (D6)	3. Date Received	3. Text Type Code	
4. Report Source Code	49. Manufacturer Contact Zip Code Ext (G1)	4. Date Removed Flag (D7)	4. Treatment	4. Patient Sequence Number	
5. Manufacturer Link Flag	50. Manufacturer Contact Country Code	5. Device Sequence No	5. Outcome	5. Date Report	
6. Number Devices in Event	51. Manufacturer Contact Postal Code	6. Date Received	6. Text (B5, or H3 or H10)		
7. Number Patient in Event	52. Manufacturer Contact Phone No Area Code (G1)	7. Brand Name (D1)			
8. Date Received	53. Manufacturer Contact Phone No Exchange (G2)	8. Generic Name (D2)			
9. Adverse Event Flag (B1)	54. Manufacturer Contact Phone No (G2)	9. Manufacturer Name (D3)			
10. Product Problem Flag (B1)	55. Manufacturer Contact Phone No Ext (G2)	10. Manufacturer Address 1 (D3)			
11. Date Report (B4)	56. Manufacturer Contact Phone No Country Code	11. Manufacturer Address 2 (D3)			
12. Date of Event (B3)	57. Manufacturer Contact Phone No City Code	12. Manufacturer City (D3)			
13. Single Use Flag (Reprocessor Flag) (D8)	58. Manufacturer Contact Phone No Local	13. Manufacturer State Code (D3)			
14. Reporter Occupation Code (E3)	59. Manufacturer GI Name (G1)	14. Manufacturer Zip Code (D3)			
15. Health Professional (E2)	60. Manufacturer GI Street 1 (G1)	15. Manufacturer Zip Code ext (D3)			
16. Initial Report to FDA (E4)	61. Manufacturer GI Street 2 (G1)	16. Manufacturer Country Code (D3)			
17. Distributor Name (F3)	62. Manufacturer GI City (G1)	17. Manufacturer Postal Code (D3)			
18. Distributor Address Line 1 (F3)	63. Manufacturer GI State Code (G1)	18. Expiration Date of Device (D4)			
19. Distributor Address Line 2 (F3)	64. Manufacturer GI Zip Code (G1)	19. Model Number (D4)			
20. Distributor City (F3)	65. Manufacturer GI Zip Code Ext (G1)	20. Lot Number (D4)			
21. Distributor State Code (F3)	66. Manufacturer GI Country Code	21. Catalog Number (D4)			
22. Distributor Zip Code (F3)	67. Manufacturer GI Postal Code	22. Other ID Number (D4)			
23. Distributor Zip Code Ext (F3)	68. Source Type (G3)	23. Device Operator (D5)			
24. Date Facility Aware (F6)	69. Date Manufacturer Received (G4)	24. Device Availability (D10)			
25. Type of Report (F7)	70. Device Date Of Manufacture (H4)	25. Date Returned to Manufacturer (D10)			
26. Report Date (F8)	71. Single Use Flag (H5)	26. Device Report Product Code			
27. Report to FDA (F11)	72. Remedial Action (H7)	27. Device Age (F9)			
28. Date Report to FDA (F11)	73. Previous Use Code (H8)	28. Device Evaluated by Manufacturer (H3)			
29. Event Location (F12)	74. Removal/Correction Number (H9)	29. Baseline Brand Name			
30. Report to Manufacturer (F13)	75. Event Type (H1)	30. Baseline Generic Name			
31. Date Report to Manufacturer (F13)		31. Baseline Model No			
32. Manufacturer Name (F14)		32. Baseline Catalog No			
33. Manufacturer Address Line 1 (F14)		33. Baseline Other ID No			
34. Manufacturer Address Line 2 (F14)		34. Baseline Device Family			
35. Manufacturer City (F14)		35. Baseline Shelf Life Contained in Label			
36. Manufacturer State Code (F14)		36. Baseline Shelf Life in Months			
37. Manufacturer Zip Code (F14)		37. Baseline PMA Flag			
38. Manufacturer Zip Code Ext (F14)		38. Baseline PMA No			
39. Manufacturer Country Code (F14)		39. Baseline 510(k) Flag			
40. Manufacturer Postal Code (F14)		40. Baseline 510(k) No			
41. Manufacturer Contact Title Name (G1)		41. Baseline Preannouncement			
42. Manufacturer Contact First Name (G1)		42. Baseline Transitional			
43. Manufacturer Contact Last Name (G1)		43. Baseline 510(k) Exempt Flag			
44. Manufacturer Contact Street 1 (G1)		44. Baseline Date First Marketed			
45. Manufacturer Contact Street 2 (G1)		45. Baseline Date Ceased Marketing			

MAUDE Files and Fields as defined on the MAUDE website. Added shading indicates file linking (see also Figure 5). Information in parentheses indicates the MedWatch Form 3500 source field.



Figure 5. MAUDE File Relationship Diagram



result in the truncation of data (at the point of the LF occurrence) within the original record, but data following the LF character may be subsequently processed as a 'new' entry, causing errors in record counts and data type incompatibilities. When importing data into SAS, this issue was circumvented by use of a TERMSTR=CRLF option in the INFILE statement.⁴²

Many files used double quotation marks to denote an exact word or phrase, e.g., THE "BATTERY VOLTAGE TOO HIGH" ALARMS WERE ACTIVATED, or to indicate inches, e.g., 7" MICROBORE NON-VENTED EXTENSION SET. And in the foidevthru1997 file, a double quotation mark was also intermittently

used to signify missing values, e.g., "|"|"|. The QUOTE option in RStudio was necessary to ensure appropriate handling of all files containing double quotation punctuation. In SAS, special handling of double quotations was only necessary when they were used to signify a missing value, in which case the TRANSTRN option can be used to identify and remove them when importing data.

Finally, given the size of the MAUDE download files, compression and keyword indexing facilitated more efficient processing in SAS. Due to the number of records in each file, it is typically impractical to review the imported data file contents in entirety; however, import problems can often be identified by:

1. Checking record counts.
2. Ensuring data types are assigned as expected.
3. Comparing the first and last imported records to the raw data files.
4. Generating descriptive summaries or tables of individual fields.

A GitHub repository with examples programs that can be used to import the MAUDE data sets into SAS is available.⁴³

Master Event Data

Data Availability

Although the majority of records are reported by the device manufacturer (Figure 1), approximately 5 percent of reports are submitted by user facilities, voluntary submitters (including health professionals or consumers/patients), and distributors. Data availability for certain fields (e.g. *DISTRIBUTOR NAME*) depends on the reporting source.

Linking to Other MAUDE Files

The Master Event file includes two database key fields: *MDR REPORT KEY* which identifies unique reports and is important for linking most of the MAUDE files together, and *EVENT KEY* which is an internally-generated field used to identify unique events. *MDR REPORT KEY* is unique within the Master Event file whereas *EVENT KEY* is not.

The MAUDE website states “A distinct master event data record will be present for each source reporting an event. In other words, if a User Facility, Distributor, Manufacturer, and voluntary submitter all report an event, there will be four event records.”²⁶ Approximately 8 percent of the Master Event file records share an *EVENT KEY*, that is, have two or more unique records present in the MAUDE database identified as pertaining to the same event. One might surmise that duplicate *EVENT KEYS* correspond to different *REPORT SOURCE CODES*

(M=manufacturer, D=distributor, U=user facility, and P=voluntary submitter). However, in practice a large majority of MAUDE reports sharing an *EVENT KEY* are submitted by the manufacturer, with many of the connected reports pertaining to devices sharing a common lot number.

Technical Tips and Considerations

An important consideration when working with the downloadable MAUDE files is whether the research question of interest pertains to unique reports or unique events. If a question involves unique events, one must incorporate the Master Event file *EVENT KEY* when preparing data for analysis, even if none of the other Master Event file data is pertinent to the question of interest.

Device Data

Data Availability

Device data is available for nearly 90 percent of MAUDE reports. Although the Device file contains baseline reporting information, the baseline reporting requirement was removed in 2008,⁴⁴ and baseline data is not supplied for most records.

Linking to Other MAUDE Files

The Device file contains an *MDR REPORT KEY* field, facilitating the linkage of this file with Master Event data. While the vast majority of the Device file records are associated with only one device (Master Event file *NUMBER DEVICES IN EVENT* = 1), there are more Device file records than MAUDE reports since a report may be associated with multiple products (Master Event file *NUMBER DEVICES IN EVENT* > 1). That is, *MDR REPORT KEY* is not unique within the Device file. This may arise, for example, if the individual submitting an adverse event report about a hip replacement system lists all of the hip system components, including the metal-on-metal hip implant, the metal liner and the femoral stem.



Figure 6. Master Event and Device File Mapping

Master Event File		Device File	
MDR REPORT KEY	NUMBER DEVICES IN EVENT	MDR REPORT KEY	DEVICE SEQUENCE NO
13190	2	13190	1
26407	1	13190	2
44272	4	26407	1
		44272	1
		44272	2
		44272	3
		44272	4

In other cases, the same problem may be linked to different instances of a medical device. For example, an infusion pump problem observed eight times, each time involving a different pump, may be submitted as one event tied to eight Device file records. The combination of *MDR REPORT KEY* and *DEVICE SEQUENCE NO* uniquely identifies a Device file record. Figure 6 illustrates the relationship between the Master Event and Device files.

Technical Tips and Considerations

Included in the Device file is a *DEVICE REPORT PRODUCT CODE* field, a unique three-letter identifier (Supplemental Table 1) indicating the type of device associated with the reported event, as determined by the submitter. There are currently over 6,000 medical device product codes, and the list is updated weekly.⁴¹ Virtually all of the Device file records include a *DEVICE REPORT PRODUCT CODE*, and this field is helpful for selecting a subset of records pertaining to a product of interest. However, readers should keep in mind that product codes are neither structured in a hierarchical manner nor validated upon entry. For example, a report about

a pacemaker lead problem may be recorded as a problem with the pacemaker itself, or even with an unrelated product.

Patient Data

Data Availability

Treatment or outcome information is present for approximately three-quarters of Patient records, though only a quarter of records include both. Patient identifying information populated in section A of the MedWatch Form 3500, including age at the time of the event, date of birth, gender and weight, is not provided in the publicly-available MAUDE database.

Linking to Other MAUDE Files

Similar to the structure of the Device file, a reported event may be linked to several patients, and there are more Patient file records than MAUDE reports. This typically occurs when a manufacturer submits one report for a problem observed in multiple individuals. In these cases, the Patient file includes $n = \text{Master Event file } \textit{NUMBER PATIENTS IN EVENT}$

records, each sharing the same *MDR REPORT KEY* but having a unique, sequentially assigned *PATIENT SEQUENCE NUMBER* ranging from 1 to n, comparable to the configuration shown in Figure 6. Consequently, the combination of *MDR REPORT KEY* and *PATIENT SEQUENCE NUMBER* uniquely identifies a Patient file record.

Technical Tips and Considerations

When using the outcome or treatment information contained in the Patient files, it is important to note that their structure differs from the instructions on the MAUDE website when multiple patients are linked to a MAUDE report.

Using the *OUTCOME* field as an example, the online MAUDE information indicates the field is structured as: "Sequence Number||','|| Outcome -- multiple source type, separate by ';'". When a single patient is linked to a report (Master Event file *NUMBER PATIENTS IN EVENT* = 1), this pattern holds true. For example, when a report is associated with one patient having outcomes indicative of a life-threatening event (*OUTCOME=L*), requiring intervention (*OUTCOME=R*) leading to hospitalization (*OUTCOME=H*), the *OUTCOME* field displays as "1,L;2,R;3,H". Enumerating a certain outcome of interest requires simply counting

mentions of a particular code or summing over indicator variables signifying the absence or presence (0/1) of an individual outcome as shown in Table 3.

However, when a MAUDE report is linked to multiple patients, the structure of the *OUTCOME* field deviates from the specified pattern in an important way – namely, the contents of the *OUTCOME* field are not specific to the patient designated by the *PATIENT SEQUENCE NUMBER*. Rather, the *OUTCOME* field represents a sequential concatenation of findings across patient records sharing a common *MDR REPORT KEY*. That is, the outcomes of each successive patient associated with an *MDR REPORT KEY* are prepended to the results from all of the previous patients having the same *MDR REPORT KEY*. This is best demonstrated through an example. Table 4a depicts a case in which a single adverse event report is associated with five patient records (Master Event file *NUMBER PATIENTS IN EVENT* = 5). As can be seen in this table, the first patient associated with the adverse event report has three outcome results: 1) hospitalization (H), 2) required intervention (R), and 3) life-threatening (L). The next patient record linked to this event (*PATIENT SEQUENCE NUMBER* = 2) contains both the *OUTCOMES* for the second

Table 3. Patient Outcome Field Structure when One Patient is Linked to a MAUDE Report

MDR REPORT KEY	PATIENT SEQUENCE NUMBER	OUTCOME	OUTCOME EVENT INDICATORS			
			HOSPITALIZATION	REQUIRED INTERVENTION	LIFE-THREATENING	DEATH
123	1	1,H;2,R;3,L;	1	1	1	0
234	1	1,D;	0	0	0	1
345	1	1,H;2,R;	1	1	0	0
456	1	1,R;	0	1	0	0
Totals:			2	3	1	1

This table shows four representative Patient file records, each associated with a different MAUDE report. The Outcome column depicts the raw data as it would be presented in the native MAUDE Patient file: H=Hospitalization; R= Required Intervention; L=Life-Threatening; D=Death. User-added 0/1 Outcome Event Indicator variables (e.g., Hospitalization, Required Intervention, Life-Threatening, Death) can facilitate record summarization as shown in the Totals row.



patient: 1) hospitalization, and those of the first patient; the third record in the series prepends the results for the next patient, and so forth.

Table 4. Patient Outcome Field Structure when Multiple Patients are Linked to a MAUDE Report

Table 4a shows the raw Patient file outcome data for an example in which five patients are associated with the same adverse event report. Color coding is used to highlight the outcomes pertaining to a specific patient: cyan for patient 1, magenta for patient 2, yellow for patient 3, gray for patient 4 and green for patient 5. In the raw data, the *OUTCOME* field for the 3rd patient associated with the adverse event report actually contains the outcomes for the 1st, 2nd, and 3rd patients. Erroneous, inflated totals for individual outcome events would be obtained if one were to naively count the mentions of each particular code as is illustrated in the user-added Hospitalization, Required Intervention, and Life-Threatening Outcome Event Indicator columns and Incorrect Totals row.

To enable correct enumeration of outcomes, Table 4b illustrates how the original *OUTCOME* field shown in Table 4a must first be deconstructed so that the parsed outcome result is specific to the patient referenced by the *PATIENT SEQUENCE NUMBER* field. Once this has been accomplished, one can correctly sum over 0/1 Outcome Event Indicator variables to obtain accurate outcome event counts as shown in the Correct Totals row.

As shown in Table 4a, simple enumeration of outcomes without taking the concatenation structure into account can result in greatly inflated totals. Because the proportion of records associated with multiple patients is quite small, failure to account for the sequential stratification of outcomes has only a minimal impact when considering the total cohort of Patient records. However, as

demonstrated in Tables 4a and 4b, the potential influence on outcome event totals can be quite sizeable if one happened to examine a small subset in which many of the adverse event reports were linked to multiple patients.

A similar situation occurs with the *TREATMENT* field. One difference with this field is that unlike the *OUTCOME* field, the *TREATMENT* field does not always contain the sequence numbers specified in the MAUDE website instructions. Instead, a simple list of treatments is typically provided, such as "COUMADIN;SYNTHROID;METFORMIN;IRON;FOSAMAX." Even when present, sequence numbers often do not follow the stated format and instead may be listed in a variety of manners, complicating parsing algorithms, e.g.: "(1) ACCU-CHEK COMPACT METER;;(2) METOPRODOL, (3) CLONIDINE;(4) NORVASC, (5)GLYBURINDE, (6) COUMADIN;(7) LEVAQUIN (8) TRAZADONE;(9) NOTROGLYCERINE;" or "1. COUMADIN;2. SYNTHROID;3. PEPSID;4. CARDIZEM;5. XYLOCAINE;6. LIDOCAINE;7. HEPARIN;". A final consideration specific to the *TREATMENT* field involves the presence of semicolons. In some cases, different treatments may be identified using the semicolon as a divider as shown in the example above. However, semicolons may not be present between treatments, e.g., "(3) CLONIDINE;(4) NORVASC, (5)GLYBURINDE, (6) COUMADIN" or may signify line breaks, e.g., "VANCOMYCIN AND GENTAMYCIN AFTER INSERTION OF NEW;GRAFT. ACE INHIBITOR: CATOPRIL 25 MG THREE TIMES;PER DAY." Parsing this last example using semicolons as the divider would lead to the unintended result of three treatments assigned to the patient: i.e., 1. VANCOMYCIN AND GENTAMYCIN AFTER INSERTION OF NEW, 2. GRAFT. ACE INHIBITOR: CATOPRIL 25 MG THREE TIMES, and 3. PER DAY.

Table 4a. Raw Outcome Data

MDR REPORT KEY	PATIENT SEQUENCE NUMBER	OUTCOME	NAÏVE OUTCOME EVENT INDICATORS		
			HOSPITALIZATION	REQUIRED INTERVENTION	LIFE-THREATENING
567	1	1,H;2,R;3,L;	1	1	1
567	2	1,H;1,H;2,R;3,L;	2	1	1
567	3	1,H;2,R;1,H;1,H;2,R;3,L;	3	2	1
567	4	1,H;2,R;1,H;2,R;1,H;1,H;2,R;3,L;	4	3	1
567	5	1,H;1,H;2,R;1,H;2,R;1,H;1,H;2,R;3,L;	5	3	1
Incorrect Totals:			15	10	5

Table 4b. Configured Outcome Data

MDR REPORT KEY	PATIENT SEQUENCE NUMBER	PARSED OUTCOME	CORRECT OUTCOME EVENT INDICATORS		
			HOSPITALIZATION	REQUIRED INTERVENTION	LIFE-THREATENING
567	1	1,H;2,R;3,L;	1	1	1
567	2	1,H;	1	0	0
567	3	1,H;2,R;	1	1	0
567	4	1,H;2,R;	1	1	0
567	5	1,H;	1	0	0
Correct Totals:			5	3	1

Based on experience, the *TREATMENT* field is best evaluated through the application of text-based tools, for example finding records that match the name of a particular medication or therapy of interest. Regardless of the approach taken to summarize treatment data, one must maintain awareness of the same sequential concatenation of results that was described for the *OUTCOME* field when multiple patients are associated with a MAUDE report.

Text Data

Data Availability

Text data is provided for the vast majority of MAUDE reports, though information may range from a single character to detailed narrative descriptions. The majority of records do not include date information.



Linking to Other MAUDE Files

Text file data is uniquely identified by the *MDR TEXT KEY* field or the combination of *MDR REPORT KEY*, *PATIENT SEQUENCE NUMBER* and *TEXT TYPE CODE*. In order to link the Text data to other files, one must utilize, the *MDR REPORT KEY* and also the *PATIENT SEQUENCE NUMBER* if linking to the Patient file.

Technical Tips and Considerations

The Text file contains more than twice the number of records present in the Master Event file. This occurs for three reasons:

1. There are $n = \text{Master Event file } \textit{NUMBER PATIENT IN EVENT}$ Text file records for each *MDR REPORT KEY*. Fortunately, unlike the Patient file data, each Text file record is specific to the referenced *PATIENT SEQUENCE NUMBER* and does not reflect a concatenation of information across patients. Fewer than 0.1% of MAUDE reports have Text file records relating to multiple patients, and the text for a given *TEXT TYPE CODE* is typically the same or very similar across patient records sharing a common *MDR REPORT KEY*.
2. *MDR REPORT KEYS* may be repeated within the Text file, reflecting the submission of supplemental or additional information for approximately 10 percent of reports.
3. The most influential factor contributing to the much larger number of Text file records is this file is populated from three separate MedWatch Form 3500 fields (B3, H3 and H10), with each comprising a separate record identified by its *TEXT TYPE CODE* (Supplemental Table 2). In practical terms, few records are associated with the H3, Manufacturer's device evaluation information, field, and information from this field is not provided in the detailed report obtained through the MAUDE online search interfaces. Even when populated, text in the H3 field

typically points back to information in the H10 field or indicates only that an investigation is pending or in process.

Our work with the MAUDE database entails text mining, and it would be highly useful to be able to easily order narrative data about a particularly product problem sequentially by time. However, *DATE REPORT* is missing for 75 percent of Text file records and when populated contains a small number of records with illogical dates extending back to 1900. Although one might assume that *MDR TEXT KEY* could be used as a surrogate for the sequential ordering of information relating to a particular report, examination of the narrative data demonstrates this, unfortunately, is not always the case (Supplemental Table 3).

Device Problems and Problem Code Descriptions

Data Availability

Approximately one-third of MAUDE reports have one or more reported *DEVICE PROBLEM CODES*. However, a quarter of these are not informative (e.g., assigned codes equating to "No Information", "Not applicable", "No code available"). A small proportion (< 10 percent) of the specified *DEVICE PROBLEM CODES* do not map to a *PROBLEM DESCRIPTION* provided in the Problem Code Descriptions file. The majority of these missing description codes may be resolved by mapping them to Component or Patient Codes available from the FDA.⁴⁵

Linking to Other MAUDE Files

Figure 7 shows the relationship between the Device Problems and Problem Code Description files. The Device Problems file may be joined to the other files using *MDR REPORT KEY*, while the Problem Code Description file must be accessed through the Device Problems file via linkage on the *DEVICE PROBLEM CODE* field.

In this example the codes associated with the first *MDR REPORT KEY* (169409) refer to an implantable cardioverter defibrillator device, and those associated to the second *MDR REPORT KEY* (184584) pertain to a saline breast prosthesis.

Technical Tips and Considerations

It may be useful to transpose the Device Problem file data from long to wide on *MDR REPORT KEY* prior to joining it with the other MAUDE files in order to link multiple problem codes with a single MAUDE report.

Similar to the discussion of the *PRODUCT CODE* field in the Device file, the *DEVICE PROBLEM CODE* field is not verified upon entry, and report submitters can mistakenly enter problem codes pertaining to products other than the one being reported.

MAUDE Data Availability Summary

The availability of data may drive or limit questions of interest. Supplemental Figure 1 provides a graphical representation of the amount of shared, non-missing data contained in the MAUDE Device, Patient, Text and Device Problems/Problem Codes

files, considering the Master Event file as the universal record. The absence of Device Problems/ Problem Code information imposes the largest constraint on data availability. Device Problem information is available for just 32 percent of records. Accordingly, if a research question depends upon coded problems, this will greatly limit the available data. In some cases, creative solutions may be useful, such as data mining the Text file information in order to identify product problems when *DEVICE PROBLEM CODEs* are not provided. In others, it may be necessary to restructure a question of interest or pursue another line of reasoning. The proportions in Supplemental Figure 1 should also be considered ‘best case’ scenarios since even though a record may exist in a file, it may not be populated with information specific to a particular inquiry. For example, while a Patient file record exists for most of the Master Event records (i.e., the same *MDR REPORT KEY* occurs in both files), outcome or treatment information is present for only 71 percent of the Patient file records and only 26 percent include both. Similarly, even among those records in which a *DEVICE PROBLEM CODE* is populated, the actual code is not informative for a sizeable proportion of reports.

Figure 7. Device Problem and Problem Code Description File Mapping

Device Problem File		Problem Code Descriptions File	
MDR REPORT KEY	DEVICE PROBLEM CODE	DEVICE PROBLEM CODE	PROBLEM DESCRIPTION
169409	1081	1081	Failure to capture
169409	1291	1291	High impedance
169409	1558	1558	Sensing intermittently
169409	1633	1633	Loss of threshold
184584	1267	1267	Gel leak
184584	1546	1546	Material rupture



Lessons Learned Summary

Table 5 summarizes some of the key lessons learned while working with the downloadable MAUDE data files, categorized by area.

Many research inquiries, including our own, require the use of two or more of the individual MAUDE files. In order to combine files appropriately, one must appreciate the data structures underlying the files. The information provided on the MAUDE website is sometimes incomplete or inaccurate, making this process difficult; however, individuals

who proceed without understanding the structural nuances of the different MAUDE file types could easily obtain misleading results. For example, failure to incorporate the Master Event file *EVENT KEY* information, even if the main parameters of interest are contained solely in the Device, Patient, Text or Device Problems files, could result in reporting of the count of unique MAUDE reports, rather than unique events. Similarly, without appreciating the structure of the data files, it can be easy to miscount records as there are more than twice as many Text file observations as MAUDE reports, and perhaps less obvious because the raw numbers are closer,

Table 5. Key Lessons and Tips for Success When Working with MAUDE Data Files

ISSUE		SUGGESTION
1. Access	The MAUDE search engines are convenient, but restrict records by time and count and provide only a small subset of available fields.	Use the downloadable MAUDE files to avoid these constraints.
2. Import	Not all of the MAUDE files include a header row containing field names.	When importing data, make sure your import process starts on the correct line.
3. Import	Some records contain non-printing characters or utilize double quotes to indicate missing information.	Account for these idiosyncrasies in the file import process to avoid errors.
4. Data Availability	MAUDE files or fields may not be populated for some/many reports.	Consider data availability when contemplating research studies using MAUDE data.
5. Linking	<i>MDR REPORT KEY</i> alone may not be sufficient when linking files.	Incorporate additional key fields when linking Patient, Device or Text file information.
6. Adverse Even Enumeration	Consider the desired denominator (reports or events) when combining files.	The <i>EVENT KEY</i> field in the Master Event file must be incorporated if enumerating unique events.
7. Field Structure	Don't assume the data will always match the format prescribed on the MAUDE website.	In particular, Patient <i>OUTCOME</i> and <i>TREATMENT</i> fields must be handled with care.

there are more Patient and Device file records than MAUDE reports since a report may be associated with multiple patients or products. If not recognized, these and other issues have the potential to bias conclusions drawn from the MAUDE database.

Conclusion

MAUDE is a valuable database for anyone interested in exploring questions about medical device safety. While the MAUDE search interfaces are useful for extracting records specific to very targeted queries, more complex questions necessitate the use of comprehensive data files provided by the FDA. When we first began working with the MAUDE files, we greatly underestimated the time required to understand the structural complexities of the various MAUDE file types. However, through this work we developed and documented a much deeper understanding of the MAUDE data and its limitations. To our knowledge, this is the first paper to detail the challenges of, and provide solutions for, using the MAUDE database files. The information in this paper would have greatly expedited our use of this resource, and we hope the lessons we learned along the way will prove useful to others and allow more effective use of MAUDE data.

Acknowledgements

Supported by NIH/NCATS Colorado CTSI Grant Number UL1 TR001082. Contents are the authors' sole responsibility and do not necessarily represent official NIH views.

References

- Szlezak N, Evers M, Wang J, Perez L. The role of big data and advanced analytics in drug discovery, development, and commercialization. *Clin Pharmacol Ther.* 2014;95(5):492-5.
- Murdoch TB, Detsky AS. The inevitable application of big data to health care. *JAMA.* 2013;309(13):1351-2.
- Costa FF. Big data in biomedicine. *Drug Discov Today.* 2014;19(4):433-40.
- Dowdy JC, Sayre RM, Shepherd JG. Indoor tanning injuries: an evaluation of FDA adverse event reporting data. *Photodermatol Photoimmunol Photomed.* 2009;25(4):216-20.
- Wang HE, Weaver MD, Abo BN, Kaliappan R, Fairbanks RJ. Ambulance stretcher adverse events. *Qual Saf Health Care.* 2009;18(3):213-6.
- Milad MP, Milad EA. Laparoscopic morcellator-related complications. *J Minim Invasive Gynecol.* 2013.
- Della Badia C, Nyirjesy P, Atogho A. Endometrial ablation devices: review of a manufacturer and user facility device experience database. *J Minim Invasive Gynecol.* 2007;14(4):436-41.
- Gurtcheff SE, Sharp HT. Complications associated with global endometrial ablation: the utility of the MAUDE database. *Obstet Gynecol.* 2003;102(6):1278-82.
- Johnson SM, Itoga N, Garnett GM, Kilcommons M, Puapong DP, Woo RK. Increased risk of cardiovascular perforation during ECMO with a bicaval, wire-reinforced cannula. *J Pediatr Surg.* 2014;49(1):46-9; discussion 9-50.
- Hauser RG, Abdelhadi R, McGriff D, Retel LK. Deaths caused by the failure of Riata and Riata ST implantable cardioverter-defibrillator leads. *Heart Rhythm.* 2012;9(8):1227-35.
- Hauser RG, Hayes DL, Almqvist AK, Epstein AE, Parsonnet V, Tyers GF, et al. Unexpected ICD pulse generator failure due to electronic circuit damage caused by electrical overstress. *Pacing Clin Electrophysiol.* 2001;24(7):1046-54.
- Hauser RG, Kallinen Retel LM. Early fatigue fractures in the IS-1 connector leg of a small-diameter ICD lead: value of returned product analysis for improving device safety. *Heart Rhythm.* 2013;10(10):1462-8.
- Andreoli JM, Lewandowski RJ, Vogelzang RL, Ryu RK. Comparison of complication rates associated with permanent and retrievable inferior vena cava filters: a review of the MAUDE database. *J Vasc Interv Radiol.* 2014;25(8):1181-5.
- Chai KE, Anthony S, Coiera E, Magrabi F. Using statistical text classification to identify health information technology incidents. *J Am Med Inform Assoc.* 2013;20(5):980-5.
- Clark KK, Sharma DK, Chute CG, Tao C. Application of a temporal reasoning framework tool in analysis of medical device adverse events. *AMIA Annu Symp Proc.* 2011:1366-71.
- Zellner C, Ports TA, Yeghiazarians Y, Boyle AJ. Sterile radial artery granuloma after transradial procedures: a unique and avoidable complication. *Catheter Cardiovasc Interv.* 2010;76(5):673-6.
- Magrabi F, Ong MS, Runciman W, Coiera E. Using FDA reports to inform a classification for health information technology safety problems. *J Am Med Inform Assoc.* 2012;19(1):45-53.
- Ross MG, Fresquez M, El-Haddad MA. Impact of FDA advisory on reported vacuum-assisted delivery and morbidity. *J Matern Fetal Med.* 2000;9(6):321-6.
- Fuller J, Ashar BS, Carey-Corrado J. Trocar-associated injuries and fatalities: an analysis of 1399 reports to the FDA. *J Minim Invasive Gynecol.* 2005;12(4):302-7.
- Institute of Medicine. Medical devices and the public health: the FDA 510(k) clearance process at 35 years. Washington, DC: National Academies Press; 2011.



21. Grigoriev I, Zu Castell W, Tsvetkov P, Antonov AV. AERS spider: an online interactive tool to mine statistical associations in Adverse Event Reporting System. *Pharmacoepidemiol Drug Saf.* 2014.
22. U.S. Food and Drug Administration. Is the product a medical device? [Accessed: 2015 September 11]. Available from: <http://www.fda.gov/medicaldevices/deviceregulationandguidance/overview/classifyyourdevice/ucm051512.htm>.
23. U.S. Food and Drug Administration. Medical Device Reporting (MDR) [Accessed: 2016 August 24]. Available from: <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/ucm2005291.htm>.
24. U.S. Food and Drug Administration. Medical Device Reporting - Alternative Summary Reporting (ASR) Program [Accessed: 2016 August 24]. Available from: <http://www.fda.gov/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm072029.htm>.
25. Safe Medical Devices Act of 1990, Federal Food, Drug, and Cosmetic Act, Public Law 101-629.
26. U.S. Food and Drug Administration. Manufacturer and user facility device experience database (MAUDE) [Accessed: 2015 December 1]. Available from: <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/PostmarketRequirements/ReportingAdverseEvents/ucm127891.htm>.
27. Mehran R, Leon MB, Feigal DA, Jefferys D, Simons M, Chronos N, et al. Post-market approval surveillance: a call for a more integrated and comprehensive approach. *Circulation.* 2004;109(25):3073-7.
28. U.S. Food and Drug Administration Center for Devices and Radiological Health. Ensuring the safety of marketed medical devices: CDRH's medical device postmarket safety program: January 2006 [Accessed: 2015 March 17]. Available from: <https://premierinc.com/about/advocacy/issues/10/udi/UDI-cdrh-report.pdf>.
29. Gurtcheff SE. Introduction to the MAUDE database. *Clin Obstet Gynecol.* 2008;51(1):120-3.
30. Latuska RF, Carlson ML, Neff BA, Driscoll CL, Wanna GB, Haynes DS. Auricular burns associated with operating microscope use during otologic surgery. *Otol Neurotol.* 2014;35(2):227-33.
31. Connell SS, Balkany TJ, Hodges AV, Telischi FF, Angeli SI, Eshraghi AA. Electrode migration after cochlear implantation. *Otol Neurotol.* 2008;29(2):156-9.
32. Erekson EA, Sung VW, Rardin CR, Myers DL. Ethylene vinyl alcohol copolymer erosions after use as a urethral bulking agent. *Obstet Gynecol.* 2007;109(2 Pt2):490-2.
33. Blumenthal KB, Sutherland DE, Wagner KR, Frazier HA, Engel JD. Bladder neck contractures related to the use of Hem-o-lok clips in robot-assisted laparoscopic radical prostatectomy. *Urology.* 2008;72(1):158-61.
34. Contractor S, Esmaeili A, Reina D, Deitch E. Incomplete deployment of the Vena Tech LP filter--case series and concerns. *Vasc Endovascular Surg.* 2011;45(4):345-51.
35. Hauser RG, Abdelhadi RH, McGriff DM, Kallinen Retel L. Failure of a novel silicone-polyurethane copolymer (Optim™) to prevent implantable cardioverter-defibrillator lead insulation abrasions. *Europace.* 2013;15(2):278-83.
36. Shah AD, Hirsh DS, Langberg JJ. Sudden and Fatal Malfunction of a Durata Defibrillator Lead due to External Insulation Failure. *Pacing Clin Electrophysiol.* 2016;39(1):101-4.
37. Mann IE, Segal OR. A case of Durata ICD lead coil externalization: Inside-out lead abrasion? *HeartRhythm Case Reports.* 2016;2(4):283-5.
38. Ensign LG. Using text mining of FDA reports to inform early signal detection of cardiovascular lead recalls: University of Colorado; 2016.
39. U.S. Food and Drug Administration. MAUDE - manufacturer and user facility device experience database search (simple, single-parameter). [Accessed: 2016 January 15]. Available from: <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm>.
40. U.S. Food and Drug Administration. MAUDE - manufacturer and user facility device experience database search (advanced, multi-parameter) [Accessed: 2016 January 15]. Available from: <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/Search.cfm>.
41. U.S. Food and Drug Administration. Download product code classification files [Accessed: 2016 January 5]. Available from: <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Overview/ClassifyYourDevice/ucm051668.htm#urc>.
42. SAS. Reading delimited text files into SAS®9, TS-673 [Accessed: 2015 November 29]. Available from: <http://support.sas.com/techsup/technote/ts673.pdf>.
43. Import MAUDE Files into SAS GitHub Respository [Internet]. [cited December 21, 2015]. Available from: <https://github.com/lensign/Import-MAUDE-Files>.
44. Medical Devices; Medical Device Reporting; Baseline Reports, 21 Code of Federal Regulations Part 803 (Effective 2008).
45. U.S. Food and Drug Administration. Coding tools / resource files [Accessed: 2015 March 7]. Available from: <http://www.fda.gov/medicaldevices/deviceregulationandguidance/postmarketrequirements/reportingadverseevents/eventproblemcodes/ucm134751.htm>.

Supplemental Table 1. FDA Medical Device Product Code Examples

PRODUCT CODE	DEVICE NAME
DTB	Permanent Pacemaker Electrode
DTC	Analyzer, Pacemaker Generator Function
DTD	Pacemaker Lead Adaptor
DTE	Pulse-Generator, Pacemaker, External
DTF	Tools, Pacemaker Service
DTG	Magnet, Test, Pacemaker
DTI	Sizer, Heart-Valve, Prosthesis
DTJ	Holder, Heart-Valve, Prosthesis

Supplemental Table 2. Types of Text Information Included in the MAUDE Text File

TEXT TYPE CODE	MEDWATCH FORM 3500 FIELD	PERCENTAGE (THRU 2013)
D	B5 = Adverse event description	50%
E	H3 = Manufacturer's device evaluation information	<1%
N	H10 = Manufacturer's additional adverse event narrative	49%




Supplemental Table 3. MDR Text Key Does Not Provide Time Sequential Ordering

MDR REPORT KEY	MDR TEXT KEY	DATE REPORT	TEXT TYPE CODE	TEXT
2280212	11975421	Missing	D	ADDITIONAL INFORMATION INDICATES THAT ADDITIONAL PRODUCT TROUBLESHOOTING WAS PERFORMED DURING A REVISION PROCEDURE AND A 1.1 JOULE COMMANDED SHOCK WAS DELIVERED AND AN OPEN CIRCUIT ERROR MESSAGE WAS DISPLAYED. SUBSEQUENTLY, THIS PATIENT'S RV LEAD WAS SURGICALLY CAPPED AND THIS CRT-D WAS EXPLANTED AND REPLACED DUE TO THE REPORTED PRODUCT PERFORMANCE ISSUE.
2280212	12745131	Missing	D	BOSTON SCIENTIFIC RECEIVED INFORMATION THAT THIS RIGHT VENTRICULAR (RV) LEAD HAD EXHIBITED AN ACUTE INCREASE IN SHOCKING IMPEDANCES GREATER THAN 125 OHMS. THE SYSTEM WAS TESTED IN ALL CONFIGURATIONS AND ALL MEASUREMENTS WERE OUT OF RANGE. THERE WAS NO NOISE ON THE SHOCK EGRAM AND THERE WERE NO EVENTS WITH THERAPY IN THE LOGBOOK SINCE IMPLANT. PRIOR TO THE THIS FINDING THE IMPEDANCES HAD BEEN CONSISTENTLY IN THE 40'S. BOSTON SCIENTIFIC TECHNICAL SERVICES (TS) DISCUSSED TROUBLESHOOTING. NO ADVERSE PATIENT EFFECTS HAVE BEEN REPORTED.
2280212	13467791	Missing	D	ADDITIONAL INFORMATION INDICATES THAT AN X-RAY WAS PERFORMED AND NO FRACTURE WAS OBSERVED. THE CAUSE FOR THE HIGH IMPEDANCE MEASUREMENTS ARE UNKNOWN AT THIS TIME. THE PATIENT WAS TO UNDERGO A REPLACEMENT PROCEDURE IN THE FUTURE.

This example illustrates that MDR Text Key is not a viable candidate for sequentially ordering in time narrative information received about a product problem.

Supplemental Figure 1. Data Shared across the MAUDE File Types



This figure represents 99.3% of the Master Event file records and includes those categories representing 1% or more of the Master Event file records thru 2013. For example, a query requiring data from the Device file could encompass up to 88% of MAUDE records (All: 28%, All but Device Problems/Problem Codes: 42%, Master Event + Device + Text: 13%, All but Patient: 3%, All but Text: 1%), while one based on Device Problems/Problem codes would be limited to 28%.